


Standard Sieve




This is standard sieve for specific surface fineness test.



Model	QRSS-901	QRSS-902
Size	Ø200 mm	Ø50 × 75 mm
Sieve Number	#170 (88 µm)	#325 (44 µm)
Weight (g)	100	30

Mortar Flow Table




This is tester used to decide mortar consistency to make sample for hydraulic cement and mortar strength test.




Model	QRMF-901
Table Diameter	Ø254 mm
Drop Height	12.7 mm
Mold	70 × 100 × 50 mm
Tamping Rod	Ø20 × 200 mm
Weight (kg)	5

Motor Type Mortar Flow Table



This is tester used to decide mortar consistency to make sample for hydraulic cement and mortar strength test.



Model	QRMF-902
Type	Motor
Falling Speed	100 times/min
Table Diameter	Ø254 mm
Drop Height	12.7 mm
Mold	70 × 100 × 50 mm
Tamping Rod	Ø20 × 200 mm
Weight (kg)	12