

## Spray Nozzle Device

This nozzle is used for wet type specific surface fineness test by standard sieve #325.

Model	QRSND-901
Nozzle Diameter	Ø17.5 mm
Nozzle Holes	Ø0.5 mm 17 ea
Material of Nozzle	BS
Pressure Gauge	2 kg/cm <sup>2</sup>
Weight (g)	2.5



## Vica Pin Apparatus

This pin is used to measure cement normal consistency and setting time of cement.

Model	QRVP-901
Needle Dia	Ø1 mm
Weight of Movable Rod	300 g
Dia Plunger	Ø10 mm
Conical Ring	60 × 70 × 40 mm
Glass Plate	100 × 100 mm
Graduated Scale	50 × 1 mm
Weight (kg)	1



## Gill more Pin Apparatus

This is equipment to measures setting time of the cements.

Model	QRGM-901
Initial Needle Dia	Ø2.12 mm
Initial Weight	113.4 g
Final Dia Plunger	Ø1.06 mm
Final Weight	453.6 g
Glass Plate	100 × 100 mm
Weight (kg)	1

